

EUROPEAN SPALLATION SOURCE

Instrumentation

J. G. Weisend II

www.europeanspallationsource.se Summer 2021







- Describe measurements & instrumentation in cryogenic systems
- Give examples of typical temperature, pressure, flow and level sensors used in cryogenic systems
- Discuss the proper installation of sensors, wiring and feed throughs in cryogenic instrumentation
- Describe Thermal Acoustic Oscillations



Introduction



- The correct measurement of properties such as temperature, pressure, flow, level and vacuum in cryogenic systems is a key factor in the success of cryogenic systems.
- Measurements will allow us to understand if our cryogenic components are working properly, enable us to control them and permit the collection of scientific data.
- There are many subtleties in the selection and installation of cryogenic sensors.
 - Poor sensor selection and installation can result in wildly inaccurate readings or sensor failure
- Think about instrumentation as a complete system sensor, wiring, feed through, DAQ rather than just the sensor itself.
 - Total system cost per measuring point can be ~ \$500 \$1000



Introduction



- Define system & sensor requirements:
 - Range
 - Accuracy
 - Time response
 - Sensor environment (e.g. presence of magnetic or radiation fields)
 - Precision what is the smallest change detected by the sensor?
 - Reliability
 - Cost



Instrumentation Rules of Thumb



EUROPEAN SPALLATION SOURCE

- Don't use more accuracy & precision than required
- Use commercially produced sensors whenever possible there is a lot available
- When possible, mount sensors outside cryostat at 300 K (e.g. pressure transducers, flow meters)
- For critical devices inside of cryostats, install redundant sensors whenever feasible
- Be sure to consider how to recalibrate sensors
- If at all possible avoid, cold instrumentation feed throughs
 SNS experience
- Once R&D is done, minimize number of sensors in series production of cryostats



Temperature Measurement



- Measure some property typically resistance or voltage drop that changes with temperature
- Commercial Temperature Sensing Options
 - Silicon Diodes (300 K ~1 K)
 - Pt Resistors (300 K ~ 30 K)
 - Ge Resistors (100 K < 1 K)
 - Carbon Glass resistors (300 K ~ 1 K) best below 100 K
 - Ruthenium Oxide (40 K < 1 K)
 - Cernox (300 K < 1 K)</p>
 - Thermocouples
- Take care not to put so much power in the sensor that it "self heats" and gives a false reading. Follow the vendor's recomendations



Typical Cryogenic Temperature Sensor Characteristics

(from LakeShore Cryotronics)



EUROPEAN SPALLATION SOURCE





Temperature Sensor Overview (from Handbook of Cryogenic Engineering)



EUROPEAN **SPALLATION** SOURCE

Table 4-6 Overview of cryogenic temperature sensors

Sensor	Measurement technique	Range [K]	Sensitivity	Stability	Size	Magneto- resistance	Radiation effect	Cost [\$]
Carbon	Resistance	0.01-300	Good	Poor	Moderate	Moderate	_	0.1
Carbon-glass	Resistance	1.4-325	Very high	Moderate	Moderate	Moderate	_	195
Capacitance	Capacitance	0.2-250	Moderate	Poor	Moderate	None	_	300
Cernox	Resistance	0.3-325	Good	Good	Small to moderate	Small	Low	125
CLTS	Resistance	4-300	Very low	Good	Large	-	Small	
CMN	Susceptibility	0.001-10	-	_		< 0.02 T	-	DIV
GaAs or GaAlAs diode	Voltage	1.4-475	Low	Good	Moderate	Moderate	-	-
Germanium	Resistance	0.05-100	Good to low	Very good	Moderate	Large	-	150 2000
³ He melting curve	Pressure	0.001-0.32	_	-	-	Small	_	130-2000 DIV
Mössbauer	Gamma detector	0.002-0.02	-	-	_	oman	-	DIV
NMR	NMR	$\mu K-mK$	-	Moderate	Very large	Moderate	_	DIV
Noise	Voltage (SQUID)	µK-300	_	Moderate	icij iugo	Moderate	-	DIV
Nuclear orientation	Gamma detector	0.004-4		Moderate	-	Small	_	680
Platinum	Resistance	10-800	Low to good	Very good	Moderate	Large	Small	75
Rhodium-iron	Resistance	0.1-600	Low to good	Very good	Small to large	Large	Small	360
Ruthenium oxide	Resistance	0.05-20	Good to low	Moderate	Moderate	Small	Sman	300
Si diode	Voltage	1.4-475	Low	Moderate	Moderate	Very large	Large	100
Superconducting fixed points	Susceptibility	0.015-7	-	Very good	Moderate	Zero field required	-	3500
Thermistor	Resistance	77-300	Very high	Good	Small	Small	-	
Thermocouple, Au-Fe	Voltage	2-300	Low	Moderate	Small	Moderate	-	10

Note: DIY in the cost column stands for Do It Yourself and can be quite expensivere 14 | Cryogenic Instrumentation - J.

Summer 2021



Cernox Temperature Sensors LakeShore Cryotronics







Very responsive at LHe Temps Expensive

Requires individual calibration for best results Very good in ionizing radiation environments

0.1

1

10

Temperature (K)

100

500



Si410 Silicon Diode Scientific Instruments



EUROPEAN SPALLATION SOURCE





Can both be individually calibrated or used with typical curves Relatively low cost, frequently used in cryogenic plants Not suitable for radiation environments



Platinum Resistors



PT-102 0.400 in ±0.050 in [10.160 mm ±1.270 mm] 0.825 in [20.955 mm] 0.079 in [2.007 mm] Ø0.010 in platinum [Ø0.254 mm] General tolerance of ±0.010 in [±0.254 mm] unless otherwise noted

Good down to ~ 30 K Works well in ionizing radiation fields Can be calibrated with good accuracy to common calibration curves Relatively low cost Excitation is generally 1 mA DC



From Lakeshore Cryotronics Catalog





- Four wire measurements (V+, V-, I+,I-) should be used for temperature sensors to avoid impact of lead resistance on measurement'
- Wires should be in twisted pairs (V+,V-) and (I+,I-) to reduce noise pickup
- Wires will connect from 300 K to cryogenic temperatures so must be have small cross sections, and low thermal conductivity
 - 36 gage manganin wire is a frequent choice
 - see thermal conductivity integrals
 - Avoid using wires that are too fine as this will result in breakage and poor reliability
 - Heat sink wires at an intermediate temperature
- Don't over constrain the wires allow room for movement and shrinkage during cool down to avoid breakage



Heat Sinking of Wires



- Critical to the proper use of temperature sensors in vacuum spaces
 - You want to measure the temperature of the sensor not that due to heat leak down the wire
 - Small heat capacities at cryogenic temperatures means small heat leaks can easily impact sensor temperature
 - Heat sink wire at intermediate temperature and also at point where temperature is measured





Table 4-3 Wire heat-sinking lengths required to thermally anchor to a heat sink at temperature T to bring the temperature of the wire to within 1 mK of T

Material	<i>T</i> 1 [K]	<i>Т</i> s [К]	Heat-sinking length, L_2 (mm) for wire sizes					
			0.21 mm ² (24 AWG)	0.032 mm ² (32 AWG)	0.013 mm ² (36 AWG)	0.005 mm ² (40 AWG)		
Copper	300	80	160	57	33	19		
	300	4	688	233	138	80		
Phosphor-	300	80	32	11	6	4		
Bronze	300	4	38	13	7	4		
Manganin	300	80	21	4	4	2		
	300	4	20	7	4	2		
304 ss	300	80	17	6	3	2		
	300	4	14	5	3	2		

Note: Values are calculated assuming wires are in a vacuum environment, and the thermal conductivity of the adhesive is given by the fit to the thermal conductivity of GE 7031 varnish.

From "Cryogenic Instrumentation" – D.S. Holmes and S. Courts Handbook of Cryogenic Engineering

Lecture 14 | Cryogenic Instrumentation - J. G. Weisend II



Pressure Measurement



- Carry out at room temperature where possible (using a capillary tube)
- Problems with room temperature pressure measurement
 - Thermal Acoustic Oscillations
 - Time response will be too slow for high speed transients but for most operations this isn't an issues
 - High speed pressure pulse due to magnet quenching is an exception
- Some cold pressure transducers exist that solve these problems
- There are a wide range of 300 K commercial pressure transducers that exist
 - Many are based on capacitive sensors or strain gage bridges mounted on diaphragms that change shape with pressure



MKS Baratron Pressure Transducer



EUROPEAN SPALLATION SOURCE

Dimensional Drawing — Note: Unless otherwise specified, dimensions are nominal values in inches (mm referenced).







300 K operation Uses a capacitive sensor Accurate up to 0.3 % of reading

Summer 2021

Lecture 14 | Cryogenic Instrumentation - J. G. Weisend II

Slide 16



Flow Measurements



- A variety of techniques are available mostly the same ones as used in standard fluid mechanics including:
 - Venturi Meters
 - Turbine Flowmeters
 - Coriolis Flowmeters
 - Orifice plate Flowmeters
- Comments
 - Insure that the devices are calibrated for operation at the temperatures and pressures that you are expecting (use appropriate fluid properties)
 - Beware of situations that can result in unplanned two-phase flow
 - Allow sufficient length for flow straightening if required (e.g. Venturi)
 - If possible install the flow meters in the 300 K portions of the flow



Venturi Flow Meter





Note use of cold DP transducer A more common approach is to use capillary tubes and a warm transducer

From The Handbook of Cryogenic Engineering



Coriolis Flow Meter





Figure 1 Coriolis meter sensor and transmitter Figure Tested down to 1.7 K at CERN LHC Flow produces vibrations in the flow tubes that have a phase offset directly related to mass flow

Figure 2 Coriolis flow meter operation

From Development of a mass flowmeter based on the Coriolis acceleration for liquid, supercritical and superfluid helium de Jonge T. et al. <u>Adv. Cryo. Engr. Vol 39 (1994)</u>



Flow Meter Comparisons (From M. Süßer – Adv.Cryo.Engr. Vol.57B – 2010)



EUROPEAN SPALLATION SOURCE

TABLE 2: Comparison of different mass flow meters

	Venturi	Orifice	V-Cone	Coriolis
Space roquirements	Inline mounting	Inline mounting	Inline mounting	Additional space
Straight upstream longth	Effects of upstream conditions	Effects of upstream conditions ⁽²⁾	Effects of upstream conditions ^{me}	No effects of upstream conditions
Connections	Capillaries	Capillaries	Capillaries	Electrical wiring
Turn down	3-5	3 - 5	3 - 5	50 (1)
Reisability	Very high	Very high	Very high	No inform.1.5
Cost of capital ⁰⁵	High	High	High	Medium
Cost of operation	Less	High	Medium	High (!)
Magnetic field	No	No	No	Yes
Dynamic measurements	Errors due frequency amplitude	Errors due frequency amplitude	Errors due frequency amplitude	Depending on frequency
Uncertainty ⁽⁴⁾	5%	5%	5%	2%
Capillary effects	yes	yes	yes	100
Pressure overload capability	high	high	high	medium

(1): Using a device with higher nominal flow rate, the pressure drop could be reduced in compensation the turn down is reduced too and the required installation space enlarged

(2): There is no information in the literature about the long term experience at low temperature conditions

(3): Pressure transducer, differential pressure transducer and temperature measurement necessary

(4): Stationary flow conditions without any capillary effect

(5): Upstream length enlarged factor 2 compared to Venturi tabes
 (6): Upstream length reduced Lengture 14 dn Gnyogenigrinstrumentation - J.



Level Measurements



- Measuring the level of cryogenic baths is important to proper operations
- Options include:
 - Capacitance gauges (LN₂, LOX, LH₂)
 - Superconducting level probes (LHe)
 - Installing redundant or easily replaceable s/c level probes is highly recommended
 - Differential pressure techniques
 - Floats (LN₂)





Feedthroughs



- Instrumentation feedthroughs are best done at 300 K
- Avoid cryogenic instrumentation feedthroughs if at all possible
 If you can't, significant testing and validation will be needed
- In test and prototype cryostats always put in more feedthroughs (or blank flanges) than you think you'll need
- GHe at 300 K and 1 bar has poor dielectric strength use spacing of pins or potting of feedthrough if this will cause problems (e.g. in voltage taps of S/C magnets)
- Beware of possible thermal acoustic oscillations being set up in pressure taps and other sealed tubes





- Occurs in a tube that connects room temperature and cryogenic temperatures and is sealed at the room temperature end
- The thermal gradient establishes standing pressure oscillations in the fluid in the tube.
- These oscillations can cause very high pressure spikes as well significant heating at the low temperature end due to friction.
- This can cause significant problems in cryogenic systems
- Such scenarios should be avoided but the enabling geometry is fairly common in cryogenic systems
 - Valved off lines
 - Pressure taps
 - Instrumentation
- There have been studies to determine stable (non-oscillating geometries)



Stability Curves for TAOs in He



EUROPEAN SPALLATION SOURCE



Y. Gu PhD Thesis University of Colorado 1993 _{Slide 24}

Lecture 14 | Cryogenic Instrumentation - J. G. Weisend II



5

Using the Stability Curves



$$\alpha = \frac{T_H}{T_c}$$

$$\varepsilon = \frac{L_H}{L_H}$$

Where the division between L_H and L_C is the point in the tube where T = (T_H + T_C)/2

These curves are for a 1 meter long tube. To use them with other lengths use the Adjusted radius:

$$r' = \frac{r}{\sqrt{L}}$$

Stability on the left hand side of the plots is due to viscous damping and stability on the right hand side of the plots is due to inertial damping





- Add volume to the warm end
- Install a check valve between the warm and cold end (near boundary between the two) – this converts the problem to a closed cold tube with no TAOs
- Heat sink the closed end (thus changing T_H/T_C)
- Allow flow through the warm end



Commercial Sources of Cryogenic Instrumentation



EUROPEAN SPALLATION SOURCE

- Don't reinvent the wheel there is a lot already available. Catalogs can help you choose the correct sensor for your application
- Two US Sources:
 - Lakeshore Cryogenics
 - http://www.lakeshore.com/
 - Scientific Instruments
 - http://www.scientificinstruments.com/
- Possible Cold Pressure Transducers
 - <u>http://www.omega.com/</u>
 - <u>http://www.gp50.com/</u>
- Cryogenic Society of America Buyer's Guide
 - <u>http://www.cryogenicsociety.org/buyers_guide/</u>